

MINI REVIEW



Radiation safety guidelines for 3D CBCT Imaging: A dental review

Navendra Jha and Fatima Injela Khan

Department of Oral Medicine & Radiology, Institute of Dental Sciences, Bareilly, Uttarpradesh, India

ABSTRACT

The field of oral and maxillofacial radiology saw transformations when Cone-beam Computed Tomography (CBCT) technology was introduced. Due to its small size, relatively low cost, and reduced exposure to ionizing radiation, CBCT was quickly adopted in dental practices compared to medical computed tomography. Inaccurate referrals for CBCT scans are still happening due to insufficient education among dentists and specialists. Moreover, technicians might raise the amount of radiation administered in order to achieve clearer images, needlessly endangering the patient. The aim of this review is to offer both patients and dentists insight into the use of CBCT technology for 3D imaging and the importance of monitoring radiation levels during CBCT procedures.

KEYWORDS

Computed tomography
cone beam; Dental
radiation safety dosimeter;
Optimal dosage

ARTICLE HISTORY

Received 28 May 2024;
Revised 12 June 2024;
Accepted 25 June 2024

Introduction

CBCT, or cone-beam processed tomography, is regularly suggested for surveying abnormalities, particularly in the craniofacial district. Contrasted with standard CT, CBCT opens patients to less radiation [1]. Given the restrictions of 2D imaging (superimpositions, mutilations, and so on), three-layered imaging (3D) has developed to satisfy the needs of improved innovation in persistent consideration while likewise being answerable for the rise of novel therapy procedures [2]. CBCT is an urgent method in dentistry due to its many purposes and results. Notwithstanding, CBCT has its own disservices, very much like any coin has different sides. How much CBCT gear at radiology focuses, confidential dentistry workplaces, and dental organizations has developed fundamentally since its presentation in the mid twentieth hundred years. Wide field of view (FOV) picture intensifiers were utilized in a few obsolete CBCT devices [3].

Subsequently, patients were in any case exposed to less radiation from these machines than from more current CBCT gear, yet not exactly from more traditional multi slice CT [4]. In view of an extensive report led in 2019, the most well-known utilizations of dental CBCT are in the fields of maxillofacial medical procedure (41%), dentoalveolar pathology (29%), orthodontics (16%), and implantology (13%) [5].

Discussion

The predicament of whether the "as low as actually attainable" (ALARA) approach might in any case be utilized for CBCT medicine emerges in light of the fact that CBCT is a fundamental piece of dentistry [6]. Over the long run, ALARA has developed into the "as low as demonstratively OK" (ALADA) technique, which assists doctors with picking the best field of view (FOV) corresponding to the return for money invested. Radiation openness is a wellbeing concern despite the fact that there is no gamble related with Dento-maxillofacial imaging for a singular patient when considered corresponding to the enormous number of individuals going through demonstrative imaging [7]. The improvement of CBCT planned to bring down the all-out radiation dose to the patient by

supplanting clinical CT in the craniofacial district. Sadly, CBCT has begun to be utilized rather than customary radiography for bitewing, all encompassing, and periapical radiographs because of an absence of severe guidelines and a broad misconception of its use in dentistry [8].

Radiographic imaging has been one of the most broadly involved examination methods for over a long time. In spite of the fact that radiographs give important data, long haul radiation concerns are raised by the radiation openness. Ongoing exploration has exhibited that radiation openness to X-beams during craniofacial indicative imaging raises the gamble of cancer [9]. Utilizing different CBCT machines from various producers and different FOV settings, Ludlow et al. led a new report on CBCT dosimetry and found that while expanding the pillar width just builds the portion to tissues previously uncovered, expanding the FOV level brings new and possibly radiosensitive tissues into the area of direct exposure [10,11]

Ludlow et al. analyzed portion and risks in oral symptomatic imaging, with an emphasis on CBCT dosimetry, utilizing a Radiation Analogue Dosimeter (RANDO) ghost outfitted with financially accessible TLD 100 TLD chips. To show where weighted tissues in the maxillofacial and neck locale that may be altogether uncovered during maxillofacial imaging are found, the Chips were organized in 24 unique positions [6]. The viable portion estimated in all-encompassing charge-coupled gadgets was 16.1 μSv , 5.6 μSv in postero-front cephalometric Photograph Stimulable Phosphor (PSP), 5.1 μSv in horizontal cephalometric PSP, 68 μSv in New Tom 3G-Enormous FOV, and 569 μSv in CB Mercuray-"Facial" FOV [3].

Qu et al's. work utilized TLD contributes a ghost to compute the mean tissue-ingested measurements for a New/Tom 9000 CBCT scanner. There were two distinct ways that the outputs were finished: with and without thyroid collars. The 2007 ICRP proposals were utilized to work out the compelling organ dose and all out powerful portion. It was

resolved that, during the collarless CBCT examine, the thyroid and throat got viable organ dosages of 31.0 μSv and 2.4 μSv , individually. There was no recognizable reduction in the organ portion when the thyroid collars were worn freely around the neck were compared [8].

Hirsch et al. utilized a human ghost stacked with TLDs to figure the measurements retained in 16 delicate organ locales. The two three-layered Accuitomo with two conventions (foremost 464 cm sweep and front 666 cm check) and the three-layered Veraviewepocs with three conventions (front 464 cm examine, front 864 cm filter, and all-encompassing + front 464 cm examine) were the two CBCT units that were utilized with various FOVs. The same and successful doses were determined by the ICRP 2005 norms. As per his discoveries, the three-layered Accuitomo 666 cm (43.27 μSv) had the most elevated powerful portion and the three-layered Accuitomo 464 cm (20.02) had the least. The successful portion for the Veraviewepocs three-layered examine was 29.78 μSv for the all-encompassing + 464 cm check convention, 30.92 μSv for the 464 cm filter, and 39.92 μSv for the 864 cm scan [9].

For E1990 and E2007, the successful portions were as per the following: 47 μSv and 78 μSv for the entire field of view (FOV); 44 μSv and 77 μSv for the 13 cm output of the jaws; 35 μSv and 58 μSv for the regular mandible; 69 μSv and 113 μSv for the high-goal mandible; and 35 μSv and 60 μSv for the high-goal maxilla. All in all: The new age of CBCT scanner has a lower powerful portion for an identical FOV as the old age hardware, as per the ICRP 2007 tissue weighting models [12].

It is significantly more vital that everybody utilizing this innovation comprehends the reasoning behind persistent openness, how to improve the patient portion, and staff radiation security methods in light of the fact that CBCT tests frequently include higher radiation dosages than conventional analytic radiography [13]. Oral and maxillofacial radiologists should appreciate the measurements and related chance of assessments and convey this data to their patients and alluding doctors. Medical care specialists need to gauge the imaging framework's expense and chance against the indicative information's conceivable worth. It has been resolved that the radiation portion from full-field-of-view dental CBCT examines is 4-42 times more prominent than that of an all-encompassing radiograph.

Powerful dose has been estimated on a scope of x-beam units utilizing the portrayed method. One significant thought while examining measurements boundaries in CBCT tests is Field of View (FOV) size [14]. Arranging FOVs into three sizes and assessing the effect of this element as an ordinal variable is enlightening. One to some degree irregular method for separating those sizes could be: For dento-alveolar imaging, a little finder (under 10 cm) is useful; for mandibula-maxillary imaging, a medium locator (10-15 cm) is satisfactory; and for maxillofacial conclusion, a major indicator (in excess of 15 cm) is preferred [15].

Powerful measurements registered with loads from 1990 and 2007 are analyzed. A mean ascent of 71% was seen with large FOV assessments, 124% with medium FOV assessments, and 181% with little FOV assessments, as per a correlation of the extent of progress by size of FOV. It is clear from analyzing the effect of changes to the powerful portion calculation that, as a result of the ICRP 2007 suggestions, risk assessment has

developed for all FOVs [16].

We have an obligation to teach our partners about the risks of contrasting "demonstratively OK" with "totally grand" photos as mindful experts. Thus, "as low as in all actuality feasible," or "as low as symptomatically satisfactory," or ALADA, was renamed and another form of ALARA was established [17,18]. To make a demonstratively OK and interpretable picture, the proper FOV, mAs, and kVp settings, along with superior quality/high goal boundaries, ought to be chosen in view of the objective of the sweep. It is worried that increasingly more CBCT filters are being performed on children and young people since they are more delicate to radiation, particularly in the thyroid, balls, and bosom tissue, and on the grounds that the gamble of disease per Sievert ascends with age [19,20].

Conclusion

Gauging the advantages and downsides of radiographic imaging is significant. As of late, CBCT has cleared the globe in a few dental strengths. It is apparent that FOV in CT and CBCT decides radiation dosages given to patients notwithstanding openness factors. Illuminating patients and dental specialists the same about the utilization of this state-of-the-art innovation and its little effect on by and large health is fundamental.

Disclosure Statement

No potential conflict of interest was reported by the authors.

References

1. Choi E, Ford NL. Measuring absorbed dose for i-CAT CBCT examinations in child, adolescent and adult phantoms. *Dentomaxillofac Radiol.* 2015;44(6):20150018. <https://doi.org/10.1259/dmfr.20150018>
2. De Vos W, Casselman J, Swennen G. Cone-beam computerized tomography (CBCT) imaging of the oral and maxillofacial region: a systematic review of the literature. *Int J Oral Maxillofac Surg.* 2009;38(6):609-625. <https://doi.org/10.1016/j.ijom.2009.02.028>
3. Ludlow JB, Timothy R, Walker C, Hunter R, Benavides E, Samuelson DB, et al. Effective dose of dental CBCT—a meta analysis of published data and additional data for nine CBCT units. *Dentomaxillofac Radiology.* 2015;44(1):20140197. <https://doi.org/10.1259/dmfr.20140197>
4. Ludlow JB, Davies-Ludlow LE, Brooks SL, Howerton WB. Dosimetry of 3 CBCT devices for oral and maxillofacial radiology: CB Mercuray, NewTom 3G and i-CAT. *Dentomaxillofac Radiol.* 2006;35(4):219-226. <https://doi.org/10.1259/dmfr/14340323>
5. Asha ML, Chatterjee I, Patil P, Naveen S. Dosimetry in dentistry. *Indian J Dent Res.* 2015;26(2):118-125. <https://doi.org/10.4103/0970-9290.159133>
6. Roberts JA, Drage NA, Davies J, Thomas DW. Effective dose from cone beam CT examinations in dentistry. *Brit J Radiol.* 2009;82(973):35-40. <https://doi.org/10.1259/bjr/31419627>
7. Ludlow JB. Dose and risk in dental diagnostic imaging: with emphasis on dosimetry of CBCT. *Imaging Sci Dent.* 2009;39(4):175-184.
8. Qu XM, Li G, Sanderink GC, Zhang ZY, Ma XC. Dose reduction of cone beam CT scanning for the entire oral and maxillofacial regions with thyroid collars. *Dentomaxillofac Radiol.* 2012;41(5):373-378. <https://doi.org/10.1259/dmfr/30200901>
9. Hirsch E, Wolf U, Heinicke F, Silva MA. Dosimetry of the cone beam computed tomography Veraviewepocs 3D compared with the 3D Accuitomo in different fields of view. *Dentomaxillofac Radiol.* 2008;37(5):268-273. <https://doi.org/10.1259/dmfr/23424132>
10. Brown J, Jacobs R, Levring Jäghagen E, Lindh C, Baksi G, Schulze D, et al. Basic training requirements for the use of dental CBCT by

- dentists: a position paper prepared by the European Academy of DentoMaxilloFacial Radiology. *Dentomaxillofac Radiol.* 2014;43(1):20130291. <https://doi.org/10.1259/dmfr.20130291>
11. Horner K, O'Malley L, Taylor K, Glenney AM. Guidelines for clinical use of CBCT: a review. *Dentomaxillofac Radiol.* 2015;44(1):20140225. <https://doi.org/10.1259/dmfr.20140225>
 12. Ludlow JB, Ivanovic M. Comparative dosimetry of dental CBCT devices and 64-slice CT for oral and maxillofacial radiology. *Oral Surgery, Oral Medicine, Oral Pathology, Oral Radiology, and Endodontology.* 2008;106(1):106-114. <https://doi.org/10.1016/j.tripleo.2008.03.018>
 13. Center OB. NCRP: Achievements of the Past 50 Years and Addressing the Needs of the Future.
 14. Davies J, Johnson B, Drage NA. Effective doses from cone beam CT investigation of the jaws. *Dentomaxillofac Radiol.* 2012;41(1):30-36. <https://doi.org/10.1259/dmfr/30177908>
 15. Jaju PP, Jaju SP. Cone-beam computed tomography: time to move from ALARA to ALADA. *Imaging Sci Dent.* 2015;45(4):263-265. <https://doi.org/10.5624/isd.2015.45.4.263>
 16. Schilling R, Geibel MA. Assessment of the effective doses from two dental cone beam CT devices. *Dentomaxillofac Radiol.* 2013;42(5):20120273. <https://doi.org/10.1259/dmfr.20120273>
 17. Harvey S, Patel S. Guidelines and template for reporting on CBCT scans. *Br Dent J.* 2020;228(1):15-18. <https://doi.org/10.1038/s41415-019-1115-8>
 18. Walliczek-Dworschak U, Diogo I, Strack L, Mandapathil M, Teymoortash A, Werner JA, et al. Indications of cone beam CT in head and neck imaging in children. *Acta Otorhinolaryngol Ital.* 2017;37(4):270-275. <https://doi.org/10.14639/0392-100X-1219>
 19. Parveen S, Kulkarni U, Mascarenhas R, Shetty R. Awareness and practice of ethics and guidelines with cone-beam computed tomography prescription in orthodontics. *J Indian Orthod Soc.* 2019;53(1):49-56. https://doi.org/10.4103/jios.jios_134_18
 20. Venkatesh E, Elluru SV. Cone beam computed tomography: basics and applications in dentistry. *J Istanbul Univ Fac Dent.* 2017;51(3):102-121. <https://doi.org/10.17096/jiufd.00289>